



عطاء: شراء وتوريد أجهزة فحص ماكنات الكلي لكل من م. قلقيلية وم. سلفيت

الجهة المستفيدة: وزارة الصحة

عطاء	عطاء/ استدراج
2012/152	الرقم

اسم المورد	رمز الصنف	اسم الصنف	الملاحظات والمواصفات	الكمية	الوحدة	سعر الوحدة	السعر الاجمالي/ شيكل	سبب الاحالة
شركة ليكس	080020003	UNIVERSAL MEASURING DEVICE	المواصفات الفنية مرفقة في ملحق الاحالة	2	عدد	17800	35600	الارخص المطابق
مجموع الاحالة شامل كافة انواع الضرائب بما فيها ضريبة القيمة المضافة								35600

رئيس لجنة العطاءات المركزية

تعرض نتيجة الاحالة لمدة يومين اعتباراً
من 2012/10/8 الى 2012/10/10

صفحة 1 من 1

عضو
مدير
مدير
مدير

عضو
عضو
عضو



التاريخ:

الرقم:

ملحق احالة العطاء رقم 152-2012

ITEM # : 1
UNIVERSAL MEASURING DEVICE
Good brand name Manufacture in USA, EUR, or JAPAN.
Multifunction meter designed to measure the conductivity, temperature and pressure of the dialysate fluids from hemodialysis delivery systems and of the water used in hemodialysis clinics.
Application: used to test and calibrate all types of delivery systems and water purification systems.
Sensors measuring either conductivity/temperature or pressure
All measurement in one device
Large display, microprocessor and keypad
All parameters displayed on the display module screen.
Long-term stability required for the most demanding applications.
With hanger
Two inputs one for fresh dialysate and other for used
USB-Interface
Power supply 230 V / 50 Hz
Conductivity
Temperature compensation is user-selectable depending on the brand of dialysis machine
Range: Approx 0 - 250 mS/cm
High resolution
High accuracy both the high and low ranges
Temperature
Range: Approx 0-100 °C
Resolution: Approx 0.01 °C
Accuracy: Approx 25-40°C ± 0.05 °C, otherwise ± 0.1 °C
Pressure
Range: Approx -700 to + 1900 mmHg, 2,5 Bar
Resolution: Approx 0.01 mmHg
Higher accuracy
Complete with:
Closing connectors for the dialysate inputs.
All tubes and sensors to work properly
All input/output connector for all sensors, probe and tubes



التاريخ:

الرقم:

Rechargeable Battery
Carrying Case
Complete with all accessories and consumables needed to work completely as specified
Country of origin
Country of source
Date of manufacturing
New catalogue including data sheet is attached with the offer.
warranty two year from installation or 30 months from delivery at least extendable for the same down time period during warranty
Warranty for not less than 10 years for spare parts.
Prices for all mentioned accessories and spares.
Certificates (agency approvals): FDA, TUV, CSA, UTRL, CE will be taken into consideration, the offer must include the certificates for evaluation
References for the same Model in Palestine
The equipment must pass the acceptance test of MOH biomedical Engineering unit.
Service manuals original and copy (2 sets in English) are included
Operation manuals original and copy (2 sets in English) are included
Your offer will not be taken in consideration if any conflict between the catalogue/data sheet submitted and the check list
THE TOTAL PRICE IN FINANCIAL OFFER SHOULD INCLUDE ALL ABOVE
OPTIONAL TO BE PRICED SEPARATELY:-
Closing connectors for the dialysate inputs.
All tubes and sensors to work properly
All input/output connector for all sensors, probe and tubes
Rechargeable Battery
Carrying Case
U.P :
T.P:
MANUFACT:
DELIVERY Date: