



قرار إحالة قطعي

رقم 12 / 2023

صادر عن لجنة العطاءات المركزية

مناقصة رقم: MOH-GSD/MOF/2022/290 شراء وتوريد وتركيب اجهزة
المراقبة القلبية الحثيثة لصالح وزارة الصحة

اجتمعت لجنة العطاءات المركزية يوم الاثنين الموافق 2023/3/13 للبت في إحالة
المناقصة أعلاه، حيث كانت اللجنة مكونة من التالية أسماؤهم:-

	رئيس لجنة العطاءات المركزية	ناصر الخطيب
	مندوب وزارة المالية	رائد التميمي
	مندوب وزارة الاقتصاد الوطني	فادي مصلح
	مندوب وزارة الأشغال العامة والإسكان	جمال قادوس
	مندوب مؤسسة المواصفات والمقاييس	مها القاضي
	مندوب الامانة العامة لمجلس الوزراء	بسام دويكات
	مندوب وزارة الاتصالات وتكنولوجيا المعلومات	علا الاسدي
	مندوب وزارة الصحة Engineer in MOH	يوليا سكر

قامت اللجنة بالاطلاع على تقرير لجنة التقييم، وعليه قررت لجنة
العطاءات المركزية ما يلي:

- 1- الغاء البند رقم (2) نظرا لارتفاع السعر عن التكلفة التقديرية
واعادة شراؤه حسب السقوف المالية.
- 2- احالة المناقصة على شركة الشرق للتكنولوجيا العالمية والمقاولات
والتجارة العامة (ستكو) كونهم تقدموا بعطاء وحيد ومستجيب
جوهريا للشروط والمواصفات.

مصادقة وزير المالية

State Of Palestine
Ministry Of Finance
General Supplies Department



دولة فلسطين
وزارة المالية
مديرية اللوازم العامة

احالة قطعية لمناقصة رقم: MOH-GSD/MOF/2022/290

رقم قرار الاحالة: 2023/12

احالة قطعية لمناقصة رقم: MOH-GSD/MOF/2022/290 شراء وتوريد وتركيب اجهزة المراقبة القلبية الحديثة لصالح وزارة الصحة

البنود المحالة على شركة الشرق للتكنولوجيا العالمية والمقاولات والتجارة العامة (ستكو):

ITEM #1 INTRA-AORTIC BALLOON PUMP:

ملاحظات	Delivery Time	Origin	MODEL	Manufacturer	الاجمالي شيقل	سعر الوحدة شيقل	الكمية	اسم البند	الرقم
According to specs: Complete with IAB catheter size30 /QTY1, IAB catheter size40 /QTY3, IAB catheter size50 /QTY2, Refillable He Cylinder QTY2, IBP Transducer QTY30; No need for UPS	16-22 Weeks	USA	AC3 optimus	Arrow	298,596.6	298,596.6	1	1 INTRA- AORTIC BALLOON PUMP	1
298596.6	اجمالي الاحالة على شركة الشرق للتكنولوجيا العالمية والمقاولات والتجارة العامة (ستكو) بالشيقل شامل ضريبة القيمة المضافة								
298596.6	اجمالي احالة المناقصة بالشيقل شامل ضريبة القيمة المضافة								

توفير شهادة مطابقة للبند المحال من مؤسسة المواصفات والمقاييس الفلسطينية قبل التوريد.

رائد التعميم/وزارة المالية فادي مصبح /وزارة الاقتصاد جمال قابوس /وزارة الاشغال العامة والإسكان بسام دوكات/ الامانة العامة لمجلس الوزراء علا الاسدي/وزارة الاتصالات وتكنولوجيا المعلومات

مصادقة وزير المالية

ناصر الخطيب
رئيس لجنة العطاءات المركزية

عضو

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس

State Of Palestine
Ministry Of Finance
General Supplies Department



دولة فلسطين
وزارة المالية
مديرية اللوازم العامة

رقم قرار الإحالة: 2023/12

احالة قطعية لمناقصة رقم: MOH-GSD/MOF/2022/290

المواصفات الفنية للبند المحال:

ITEM #1 INTRA-AORTIC BALLOON PUMP

MANUFACTURER	Arrow
MODEL	AC3 optimus
COUNTRY OF ORIGIN	USA
WARRANTY	5 years
DELIVERY	16-22 Weeks
QTY	1
TECHNICAL SPECIFICATIONS	
Good brand name Manufacturer in (USA, EUR, or JAPAN) - Well Known brand name with good reputation	USA
Equipment should be FDA approved	YES
Latest generation IABP system	YES
Must be able to accept fiber-optic and non-fiber optic catheters	YES
Microprocessor / microcontroller based system	YES
System should be complete with Display Control system and pneumatic drive unit	YES

رائد التميمي/وزارة المالية فادي مصلح /وزارة الاقتصاد جمال قانوس / وزارة الاشغال العامة والإسكان بسام دويكات/ الامانة العامة لمجلس الوزراء علا الاسدي/وزارة الاتصالات وتكنولوجيا المعلومات

ناصر الخطيب

مصادقة وزير المالية

رئيس لجنة العطاءات المركزية

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس



System should be transportable and compact with adequate battery backup	YES
System should start up easy and user friendly	YES
System should have fast technique to provide accurate & reliable ventricular support enhancing augmentation & improved after-load reduction	YES
System should have automatic and manual operating modes, ECG and pressure triggering modes are required	YES
System should have at least 2 types of assist ratio like 1:1, 1:2 for comfortable weaning. (Frequency Weaning)	YES
System should be capable of automatically selecting appropriate trigger i.e ECG or pressure and also accurately select the inflation and deflation points, in automatic mode	YES
In automatic mode of operation user should be in control of the deflation point	YES
In automatic mode, advanced software should automatically adapt the timings for various rhythms and rate variations, without any user intervention	YES
In automatic mode, it should automatically identify atrial fibrillation & adopt R-Wave deflation mode for better patient support	YES
In manual mode, the system should allow user to control most of the pump functions	YES

رائد التميمي/وزارة المالية /فادي مصلح /وزارة الاقتصاد جمال غانوس /وزارة الأشغال العامة والإسكان بنسام دويكات/ الامانة العامة لمجلس الوزراء علا الاسدي/وزارة الإتصالات وتكنولوجيا المعلومات

Jaleel

ناصر الخطيب

مصادقة وزير المالية

رئيس لجنة العطاءات المركزية

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس



Should be able to trigger at lower pulse pressure in patients with hypotension in pressure trigger mode	YES
Should have on screen indication of Helium level in the cylinder and battery level for timely intervention and correction	YES
Should be able to display at last 3 waveform as ECG, invasive pressure and balloon pressure waveforms	YES
Large display (at least 12inch diagonal display) for brighter & very good visibility from a distance in any lighting conditions	YES
IABP monitor should display inflation and deflation intervals on the pressure waveform to allow for timing adjustments. ECG, heart rate, arterial pressure, and alarm conditions should also be on the display. Balloon pressure, volume, and setup instructions should be accessible, but not necessarily displayed at all times	YES
ECG inflation marker to indicate inflation period on ECG which can be useful when pressure waveform is not available	YES
As ECRI recommendations, IABP is life-support system; therefore, alarms should be distinct and easily identified. Alarms associated with critical parameters (e.g., trigger loss, vacuum loss, balloon disconnect, balloon overinflation / leak) are necessary and should activate both audible and visual alarms and deflate the balloon	YES

Handwritten signature

علا الاسدي/وزارة الاتصالات وتكنولوجيا المعلومات

بسام دويكات/ الامانة العامة لمجلس الوزراء

جمال قانوس / وزارة الاشغال العامة والإسكان

قاسم مصلح /وزارة الاقتصاد

رائد التميمي/وزارة المالية

ناصر الخطيب

رئيس لجنة العطاءات المركزية

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس



State Of Palestine
Ministry Of Finance
General Supplies Department

رقم قرار الإحالة: 2023/12

As ECRI recommendations , IABP should also alarm in the event of power loss, inflation during systole, and significant heart rate change. If an alarm is silenced, a visual display should clearly indicate which alarm is disabled. It should also be possible for the operator to silence certain alarms while leaving others active	YES
On screen indication of standby time and should give alarm after 20 min., to draw user's attention on the system being on standby	YES
Fiber optic pressure signal output should be available for external monitor to eliminate need for additional pressure monitoring site & transducer	YES
Should give extensive help to correct the alarm conditions that are specific to the alarm condition	YES
Should have facility for condensation removal automatically without user intervention should be maintenance free	Need to empty the bottle manual, Acceptable
Should provide peripheral vascular Doppler for checking limb ischemia	NA, Acceptable
Pump console should have a chart recorder for printing out patient hemodynamic data	YES
In-built comprehensive service diagnostics to help the technician to locate the fault immediately	YES

سورة

علا الاسدي/وزارة الاتصالات وتكنولوجيا المعلومات

بسام دويكات/ الامانة العامة لمجلس الوزراء

جمال قانوس / وزارة الاشغال العامة والإسكان

رائد التميمي/وزارة المالية فادي مصلح /وزارة الاقتصاد

ناصر الخطيب

رئيس لجنة العطاءات المركزية

مصادقة وزير المالية

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس



Should have capability to connect on the hospital network and should be provided with the necessary software with multiple concurrent user licenses	YES
Should have valid certification from the international standards organizations. certificate should be submitted	YES
System Shall meet IEC-60601-1-2:2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility. Or should comply with 89/366/EEC; EMC directive. certificate should be submitted	YES
Suitable On line UPS of suitable rating with voltage regulation and spike protection for 60 minutes back up.(Should be priced separately)	YES/INCLUDE
Price of an additional year after the five years required with the device, should be specified/mentioned in the offer	YES/OPTIONAL
System Configuration Accessories, spares and consumables: Prices of consumables must be specified in the bid and running cost will be taken into consideration during the bid evaluation process.	YES
ALL ACCESSORIES BELOW SHOULD BE PRICED	
ECG Thermal Paper, (10 Pcs/box)	YES/INCLUDE
ECG Cable with one box electrode (QTY: 2set)	YES/INCLUDE

علا الاسدي/وزارة الاتصالات وتكنولوجيا المعلومات

بسام دويكات/ الامانة العامة لمجلس الوزراء

جمال قادوس / وزارة الاشغال العامة والإسكان

رائد التميمي/وزارة المالية كادي مصلح /وزارة الاقتصاد

ناصر الخطيب

رئيس لجنة العطاءات المركزية

مصادقة وزير المالية

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس



Disposable invasive blood pressure transducer (QTY: 30)	YES/OPTIONAL
Refillable helium tank cylinder not less than 100L Helium (@2200psi) compatible with the IABP system & with suitable head to be refilled locally (QTY: 2)	YES/OPTIONAL
Disposable helium tank cylinder not less than 140L Helium (@2300psi) compatible with the IABP system (QTY: 5)	YES/INCLUDE
Intra-aortic balloon catheter, size 30cc (Qty: 2)	YES/INCLUDE
Intra-aortic balloon catheter, size 35cc (Qty: 2)	NA
Intra-aortic balloon catheter for adult, size 40cc (Qty: 2)	YES/INCLUDE
Intra-aortic balloon catheter for adult, size 50cc (Qty: 2)	YES/INCLUDE
Other accessories to be priced separately:	
Fiber-Optic Intra-aortic balloon catheter, Sizes: 30-35, 40 & 50cc	YES/OPTIONAL
DOCUMENTATION	
1. Operating Manual (in original) soft and hard copies	YES
2. Service Manual (in original).soft and hard copies include Maintenance ,Troubleshooting, Electronic boards schematics , parts list .	YES
3. List of important spare parts and accessories with their part number and costing.	YES

رائد التميمي/وزارة المالية فادي مصلح/وزارة الاقتصاد جمال قانوس/وزارة الاشغال العامة والإسكان بسام دونكات/ الامانة العامة لمجلس الوزراء علا الاسدي/وزارة الاتصالات وتكنولوجيا المعلومات

ناصر الخطيب

مصادقة وزير المالية

رئيس لجنة العطاءات المركزية

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس

State Of Palestine
Ministry Of Finance
General Supplies Department



دولة فلسطين
وزارة المالية
مديرية اللوازم العامة

رقم قرار الإحالة: 2023/12

أحالة قطعية لمنافسة رقم: MOH-GSD/MOF/2022/290

4. Certificate of calibration and inspection.	YES
5. List of Equipment available for providing calibration and routine maintenance support as per manufacturer documentation in service / technical manual.	YES
STANDERDS: Must be submitted with the offer ,with valid dates not less than 6 months	YES
1- Equipment should be FDA, European CE, UL or BIS approved product. certificate should be submitted	YES
2- CE free sale certificate should be submitted	YES
3- Manufacturer should have ISO certification for quality standards . certificate should be submitted	YES
4- Certificate from Palestenian Standard Institution	YES
TRAINING	
On-site training of 4 x operators to departmental faculty and staff for optimum usage till satisfaction.	YES
On-site training for MOH engineers for equipment troubleshooting as well as PPM	YES/INCLUDE
INSPECTION	
1. Inspection Authority: Biomedical Engineering Unit	YES
2. Shall inspect and test and where necessary reject the equipments after its arrival at the Hospital	YES

رائد التميمي/وزارة المالية فاتحي مصلح /وزارة الاقتصاد جمال قانوس / وزارة الاشغال العامة والإسكان بسام دويكات/ الامانة العامة لمجلس الوزراء علا الاسدي/وزارة الاتصالات وتكنولوجيا المعلومات

ناصر الخطيب

مصادقة وزير المالية

رئيس لجنة العطاءات المركزية

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس

State Of Palestine
Ministry Of Finance
General Supplies Department



دولة فلسطين
وزارة المالية
مديرية اللوازم العامة

رقم قرار الإحالة: 2023/12

أحالة قطعية لمنافسة رقم: MOH-GSD/MOF/2022/290

Installation	
Installation, Testing, Commissioning & Handing-over	YES
Warranty	
1. Warranty: 5 years, on site comprehensive warranty (labour & spares covering all parts of the units and items supplied), from the date of issue of installation certificate by Biomedical Engineering Unit	5 YEARS
2. 95% uptime guarantee should be given. In case downtime exceeds 5%, penalty in the form of extended warranty, double the number of days for which the equipment goes out of service will be applied.	YES
3. In case of agency removed from the agent or transferred to other company, the old agent should be responsible of all warranty Implications, such as service , spare parts ,, etc	YES
MAINTENANCE & REPAIR	
1. The supplier must ensure the availability of expertise service and maintenance The vendor should submit company profile including names and No. of engineers, training certifications preferably on the same product.	YES
2. The system should be upgradable and all spare parts should be available for next 10 years after commissioning	YES
3. Application and service softwares must be provided to the Biomedical Engineering unit free of charge / or must be on demand	YES
TERMS AND CONDITIONS	

رائد التميمي/وزارة المالية فادي مصلح /وزارة الاقتصاد جمال قادوس / وزارة الأشغال العامة والإسكان بسام دويكات/ الامانة العامة لمجلس الوزراء علا الاسدي/وزارة الاتصالات وتكنولوجيا المعلومات

مصادقة وزير المالية

رئيس لجنة العطاءات المركزية

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس

State Of Palestine
Ministry Of Finance
General Supplies Department



دولة فلسطين
وزارة المالية
مديرية اللوازم العامة

MOH-GSD/MOF/2022/290: احالة قطعية لمنافسة رقم

رقم قرار الإحالة: 2023/12

1.The Supplied equipment should be complete with all accessories and consumables needed to work completely as specified	YES
2.Country of origin should be clear in the offer	YES
3.Country of source should be clear in the offer	YES
4.Date of manufacturing should be clear in the offer: The offered model must be new , manufacturing date not more than 1 year from the tender date	YES
5. Original catalogue and detailed datasheet for the quoted model to be enclosed.	YES
6.Number of units installed in Palestine if any (a list of the same model with serial number and location).	YES
7.The offer will not be taken in consideration in case of incomplete compliance sheet or any conflict between the catalogue/data sheet submitted and the compliance sheet	YES
8.The Vendor should submit Agency Agreement, or Sole distribution agreement or Authorization letter for sale and maintenance.	YES
9. Company must be registered in MOH	YES

رائد التميمي/وزارة المالية فادي مصلح /وزارة الاقتصاد جمال قانوس /وزارة الأشغال العامة والإسكان بسام دويكات/ الامانة العامة لمجلس الوزراء علا الاسدي/وزارة الاتصالات وتكنولوجيا المعلومات

ناصر الخطيب

مصادقة وزير المالية

رئيس لجنة العطاءات المركزية

عضو

عضو

مها القاضي /مؤسسة المواصفات والمقاييس