



التاريخ: ٣١ / ٥ / ٢٠٢٤

الرقم:

تنويه صادر عن اللوازم العامة بخصوص المناقصة رقم: (MMS-GSD/MOF/2024/120)

اسم المناقصة: شراء وتوريد وتركيب أجهزة طبية لتجهيز عيادات الخليل (دورا-يطا-حلو)

لصالح مديرية الخدمات الطبية العسكرية

إشارة إلى الموضوع المذكور أعلاه، وإلى استفسارات الشركات، يرجى ملاحظة المواصفات التالية:

Item NO 2 : DIAGNOSTIC SET OPHTHALMOSCOPE / OTOSCOPE – WALL

MOUNTED:

SHOULD BE FULL SPECIFICATIONS OF WALL MOUNTED TYPE.

ITEM NO. 5 : MOBILE EXAMINATION LAMP

LED BUBLS CAN BE ACCEPTED FROM MANUFACTURER AND IT SHOULD BE LONG LIFE LED MODULE.

ITEM NO. 8 : VITAL SIGN MONITOR

- a. The monitor must include 2 pcs. of adult Finger SPO2 probes and extension cable, one standard and one spare.
- b. The monitor must include 2 pcs. of child Finger SPO2 probes and extension cable, one standard and one spare.
- c. The monitor must include 3 pcs of each adult size (large adult, adult, small adult) of NIBP cuffs, hoses and extension cables, (3 set of all size).
 1. The monitor must include just one pcs of child adult size of NIBP cuffs, hoses and extension cables.
 2. The monitor must include 500 pcs of ear speculum for each device.

يعتبر هذا التنويه جزاً لا يتجزأ من وثائق المناقصة وسيتم اعتماده عند التقييم وعلى جميع الشركات ارفاقه مع ملف العطاء.

